



Eastern Analytical, Inc.

professional laboratory and drilling services

Bob Carter
Boys & Girls Clubs of Central NH
55 Bradley Street
Concord, NH 03301



Subject: Laboratory Report

Eastern Analytical, Inc. ID: 197015
Client Identification: Concord Bradley Street Unit | CB-06411
Date Received: 6/24/2019

Dear Mr. Carter:

Enclosed please find the laboratory report for the above identified project. All analyses were performed in accordance with our QA/QC Program. Unless otherwise stated, holding times, preservation techniques, container types, and sample conditions adhered to EPA Protocol. Samples which were collected by Eastern Analytical, Inc. (EAI) were collected in accordance with approved EPA procedures. Eastern Analytical, Inc. certifies that the enclosed test results meet all requirements of NELAP and other applicable state certifications. Please refer to our website at www.easternanalytical.com for a copy of our NELAP certificate and accredited parameters.

The following standard abbreviations and conventions apply to all EAI reports:

- Solid samples are reported on a dry weight basis, unless otherwise noted
- < : "less than" followed by the reporting limit
- > : "greater than" followed by the reporting limit
- %R : % Recovery


Eastern Analytical Inc. maintains certification in the following states: Connecticut (PH-0492), Maine (NH005), Massachusetts (M-NH005), New Hampshire/NELAP (1012), Rhode Island (269), Vermont (VT1012) and New York (12072).

The following information is contained within this report: Sample Conditions summary, Analytical Results/Data, Quality Control data (if requested) and copies of the Chain of Custody. This report may not be reproduced except in full, without the the written approval of the laboratory.

If you have any questions regarding the results contained within, please feel free to directly contact me or the chemist(s) who performed the testing in question. Unless otherwise requested, we will dispose of the sample (s) 30 days from the sample receipt date.

We appreciate this opportunity to be of service and look forward to your continued patronage.

Sincerely,


Lorraine Olashaw, Lab Director

7.2.19
Date

4
of pages (excluding cover letter)



SAMPLE CONDITIONS PAGE

EAI ID#: 197015

Client: **Boys & Girls Clubs of Central NH**

Client Designation: **Concord Bradley Street Unit | CB-06411**

Temperature upon receipt (°C): **22**

Received on ice or cold packs (Yes/No): **N**

Acceptable temperature range (°C): 0-6

Lab ID	Sample ID	Date Received	Date Sampled	Sample Matrix	% Dry Weight	Exceptions/Comments (other than thermal preservation)
197015.01	Junior Game Room Bubblers	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.02	PreK Room Bubblers	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.03	Cafeteria Bubblers 1	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.04	Cafeteria Bubblers 2	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.05	Kitchen 3 Bay Sink	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.06	Kitchen Food Prep Sink	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.07	Teen Room Bubblers 1	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.08	Teen Room Bubblers 2	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy
197015.09	Senior Room Kitchen Sink	6/24/19	6/22/19	aqueous		Adheres to Sample Acceptance Policy

Samples were properly preserved and the pH measured when applicable unless otherwise noted. Analysis of solids for pH, Flashpoint, Ignitability, Paint Filter, Corrosivity, Conductivity and Specific Gravity are reported on an "as received" basis.

Immediate analyses, pH, Total Residual Chlorine, Dissolved Oxygen and Sulfite, performed at the laboratory were run outside of the recommended 15 minute hold time.

All results contained in this report relate only to the above listed samples.

References include:

- 1) EPA 600/4-79-020, 1983
- 2) Standard Methods for Examination of Water and Wastewater, 20th, 21st, 22nd & 23rd Edition or noted Revision year.
- 3) Test Methods for Evaluating Solid Waste SW 846 3rd Edition including updates IVA and IVB
- 4) Hach Water Analysis Handbook, 4th edition, 1992



LABORATORY REPORT

EAI ID#: 197015

Client: **Boys & Girls Clubs of Central NH**

Client Designation: **Concord Bradley Street Unit | CB-06411**

Sample ID:	Junior Game Room Bubbler	PreK Room Bubbler	Cafeteria Bubbler 1						
Lab Sample ID:	197015.01	197015.02	197015.03						
Matrix:	aqueous	aqueous	aqueous						
Date Sampled:	6/22/19	6/22/19	6/22/19	Analytical		Date of			
Date Received:	6/24/19	6/24/19	6/24/19	Matrix	Units	Analysis	Method	Analyst	
Lead	0.0016	< 0.001	< 0.001	AqTot	mg/L	6/27/19	200.8	HEH	

Sample ID:	Cafeteria Bubbler 2	Kitchen 3 Bay Slnk	Kitchen Food Prep Sink						
Lab Sample ID:	197015.04	197015.05	197015.06						
Matrix:	aqueous	aqueous	aqueous						
Date Sampled:	6/22/19	6/22/19	6/22/19	Analytical		Date of			
Date Received:	6/24/19	6/24/19	6/24/19	Matrix	Units	Analysis	Method	Analyst	
Lead	< 0.001	< 0.001	< 0.001	AqTot	mg/L	6/25/19	200.8	HEH	

PreK Room Bubbler and Cafeteria Bubbler 1 analyzed on 6/25/2019.



LABORATORY REPORT

EAI ID#: 197015


Client: **Boys & Girls Clubs of Central NH**

Client Designation: **Concord Bradley Street Unit | CB-06411**

Sample ID:	Teen Room Bubbler 1	Teen Room Bubbler 2	Senior Room Kitchen Sink					
Lab Sample ID:	197015.07	197015.08	197015.09					
Matrix:	aqueous	aqueous	aqueous					
Date Sampled:	6/22/19	6/22/19	6/22/19	Analytical Matrix	Units	Date of Analysis	Method	Analyst
Date Received:	6/24/19	6/24/19	6/24/19					
Lead	< 0.001	< 0.001	< 0.001	AqTot	mg/L	6/25/19	200.8	HEH

Chain of Custody Record

197015

Program Manager: Bob Carter	Page 1 of 1 
Company: Boys & Girls Clubs of Central NH	
Address: 55 Bradley Street	
City, State Zipcode: Concord, NH 03301	
Phone: 224-1061	
E-mail: carter@centralnhclubs.org; emond@centralnhclubs.org	

Sample ID's	Date/Time of Collection	Lead Analysis
Junior Game Room Bubblers	6/22/19 8:00 AM	✓
PreK Room Bubblers	6/22/19 8:00 AM	✓
Cafeteria Bubblers 1	6/22/19 8:00 AM	✓
Cafeteria Bubblers 2	6/22/19 8:00 AM	✓
Kitchen 3 Bay Sink	6/22/19 8:00 AM	✓
Kitchen Food Prep Sink	6/22/19 8:00 AM	✓
Teen Room Bubblers 1	6/22/19 8:00 AM	✓
Teen Room Bubblers 2	6/22/19 8:00 AM	✓
Senior Room Kitchen Sink	6/22/19 8:00 AM	✓
		✓
		✓
		✓
		✓
		✓
		✓
		✓

Monitoring Location Identifier: CB-06411 Permission to Upload: Yes No

Collected by: (Print) Bob Carter
 Relinquished by: [Signature] Date/Time: 6/24/19 8:45 AM
 Received by: [Signature] Date/Time: 6/24/19 8:45

Relinquished by: _____ Date/Time: _____
 Received by: _____ Date/Time: _____

Hebe vic