



Eastern Analytical, Inc.

professional laboratory and drilling services

Bob Carter
Boys & Girls Clubs of Central NH
55 Bradley Street
Concord, NH 03301



Subject: Laboratory Report

Eastern Analytical, Inc. ID: 197016
Client Identification: Suncook Unit | CB-06853
Date Received: 6/24/2019

Dear Mr. Carter :

Enclosed please find the laboratory report for the above identified project. All analyses were performed in accordance with our QA/QC Program. Unless otherwise stated, holding times, preservation techniques, container types, and sample conditions adhered to EPA Protocol. Samples which were collected by Eastern Analytical, Inc. (EAI) were collected in accordance with approved EPA procedures. Eastern Analytical, Inc. certifies that the enclosed test results meet all requirements of NELAP and other applicable state certifications. Please refer to our website at www.easternanalytical.com for a copy of our NELAP certificate and accredited parameters.

The following standard abbreviations and conventions apply to all EAI reports:

- Solid samples are reported on a dry weight basis, unless otherwise noted
- < : "less than" followed by the reporting limit
- > : "greater than" followed by the reporting limit
- %R : % Recovery

Eastern Analytical Inc. maintains certification in the following states: Connecticut (PH-0492), Maine (NH005), Massachusetts (M-NH005), New Hampshire/NELAP (1012), Rhode Island (269), Vermont (VT1012) and New York (12072).

The following information is contained within this report: Sample Conditions summary, Analytical Results/Data, Quality Control data (if requested) and copies of the Chain of Custody. This report may not be reproduced except in full, without the the written approval of the laboratory.

If you have any questions regarding the results contained within, please feel free to directly contact me or the chemist(s) who performed the testing in question. Unless otherwise requested, we will dispose of the sample (s) 30 days from the sample receipt date.

We appreciate this opportunity to be of service and look forward to your continued patronage.

Sincerely,

Lorraine Olashaw
Lorraine Olashaw, Lab Director

7.2.19
Date

3
of pages (excluding cover letter)



SAMPLE CONDITIONS PAGE

EAI ID#: 197016

Client: **Boys & Girls Clubs of Central NH**

Client Designation: **Suncook Unit | CB-06853**

Temperature upon receipt (°C): 22

Received on ice or cold packs (Yes/No): N

Acceptable temperature range (°C): 0-6

Lab ID	Sample ID	Date	Date	Sample	% Dry	Exceptions/Comments (other than thermal preservation)
		Received	Sampled	Matrix	Weight	
197016.01	Kitchen 3 Bay Sink	6/24/19	6/21/19	aqueous		Adheres to Sample Acceptance Policy
197016.02	Kitchen Food Prep Sink	6/24/19	6/21/19	aqueous		Adheres to Sample Acceptance Policy
197016.03	Downstairs Water Bubblers	6/24/19	6/21/19	aqueous		Adheres to Sample Acceptance Policy

Samples were properly preserved and the pH measured when applicable unless otherwise noted. Analysis of solids for pH, Flashpoint, Ignitability, Paint Filter, Corrosivity, Conductivity and Specific Gravity are reported on an "as received" basis.

Immediate analyses, pH, Total Residual Chlorine, Dissolved Oxygen and Sulfite, performed at the laboratory were run outside of the recommended 15 minute hold time.

All results contained in this report relate only to the above listed samples.

References include:

- 1) EPA 600/4-79-020, 1983
- 2) Standard Methods for Examination of Water and Wastewater, 20th, 21st, 22nd & 23rd Edition or noted Revision year.
- 3) Test Methods for Evaluating Solid Waste SW 846 3rd Edition including updates IVA and IVB
- 4) Hach Water Analysis Handbook, 4th edition, 1992



LABORATORY REPORT

EAI ID#: 197016


Client: **Boys & Girls Clubs of Central NH**

Client Designation: **Suncook Unit | CB-06853**

Sample ID:	Kitchen 3 Bay Sink	Kitchen Food Prep Sink	Downstairs Water Bubbler					
Lab Sample ID:	197016.01	197016.02	197016.03					
Matrix:	aqueous	aqueous	aqueous					
Date Sampled:	6/21/19	6/21/19	6/21/19	Analytical Matrix	Units	Date of Analysis	Method	Analyst
Date Received:	6/24/19	6/24/19	6/24/19					
Lead	< 0.001	< 0.001	< 0.001	AqTot	mg/L	6/25/19	200.8	HEH



Chain of Custody Record

197016

Program Manager: Bob Carter	Page 1 of 1 
Company: Boys & Girls Clubs of Central NH	
Address: 55 Bradley Street	
City, State Zipcode: Concord, NH 03301	
Phone: 224-1061	
E-mail: carter@centralnhclubs.org; emond@centralnhclubs.org	

Project Name: Suncook Unit		
Sample ID's	Date/Time of Collection	Lead Analysis
Kitchen 3 Bay Sink	6/21/19 6:45 AM	√
Kitchen Food Prep Sink	6/21/19 6:45 AM	√
Downstairs Water Bubbler	6/21/19 6:45 AM	√
		√
		√
		√
		√
		√
		√
		√
		√
		√
		√
		√
		√

Monitoring Location Identifier: CB-06853 Permission to Upload: Yes No

Collected by: (Print): Bob Carter
 Relinquished by:  Date/Time: 6/24/19 8:45 AM
 Received by:  Date/Time: 6/24/19 8:45

Relinquished by: _____ Date/Time: _____
 Received by: _____ Date/Time: _____

No be, 20°C